

CONFIDENTIAL

31 Jan 61 25X1

ENG M-61-

Chief, Operations Division, Medical Staff

Chief, Engineering Staff, OC

Status of Radiological Survey and Radiation Detection
Instruments [redacted] Quarterly Report

To be completed

25X1

1. During October, November and December 1960, the radiological survey and radiation detection equipments [redacted] were tested to determine their operational status. Three instruments were found inoperative because of dead batteries.

25X1
25X1

2. Replacement batteries for all units will be ordered in January 1961. Calibration of these instruments will be made after receipt of fresh batteries.

3. Four man-hours were expended in maintenance and repair during the last three months. Total man-hours for 1960: 45.

25X1

R&D/Lab/JWM/rkb (31 January 1961)

Distribution:

- Original and 1 - Addressee
- 1 - Lab Subj.
- 1 - OC-E Chrono
- 1 - Dev/s
- 1 - Lab Pending ←

DOC	9	REV DATE	5-12-80	BY	008632
ORIG COMP		OR			C-
ORIG CLASS	C	DATE	1		
JUST	22	NEXT	2010		

CONFIDENTIAL

~~SECRET~~

18 October 1960

MEMORANDUM FOR: Chief, Engineering Staff, OC

ATTENTION :

FROM :

SUBJECT :

Administrative Officer, Operations Division, MS

Quarterly Report of Status of Radiological Survey
and Radiation Detection Instruments

REFERENCE :

ENG M 60-818, 10 October 1960

1. This is to inform you that [] is no longer
the Administrative Officer, Operations Division, Medical Staff.

2. In the future if you will please send the reports to the
AO/OD/MS without using any name, the reports will get the proper
attention.

[]
Administrative Officer
Operations Division, Medical Staff

DOC	10	REV DATE	5-12-80	BY	008632
ORIG COMP		ORIG			
ORIG CLASS	5	PROJ	1		
JUST	0.2	NEXT REV	2010		

CONFIDENTIAL